
May 6, 2010



Product Safety Alert

UL warns of dental endodontic device

Northbrook, Ill., USA - Underwriters Laboratories (UL) is notifying dentists, retailers, distributors, and consumers that the product, identified below, bears a label containing a mark resembling UL's certification mark for the United States and Canada. This product has not been evaluated by UL to the appropriate standard for safety for the United States or Canada, and is not authorized to bear the UL Mark. It is unknown if these products comply with United States or Canadian safety requirements.

Name of Product: Endomax Smart

Number of Units: Unknown Quantity

Brand Name: Maximum Dental

Date of Manufacture: Unknown

Product Identification / On the Product:

ENDOMAX SMART Canal Treatment Device
Model EMS
Type EL
Volts 4.5V-/9.6V-
Amps 0.03A
Internally Powered Equipment
SER No. DMEN00295

Manufactured by MAXIMUM DENTAL
Secaucus New Jersey-USA

Known Place of Sale:

Maximum Dental
600 Meadowland Pkwy #269
Secaucus, New Jersey 07094
Phone: 201-617-1101

May 6, 2010

