

May 5, 2011

Eppendorf North America Recalls MixMate Laboratory Plate and Tube Mixer Due to a Potential Burn Risk When Moisture Enters the Housing

The Electrical Safety Authority is notifying the public that Eppendorf North America has announced a voluntary recall of the following consumer product. Consumers should stop using recalled products immediately.

Name of Product: Eppendorf MixMate

Units: 26 units sold in Ontario and 61 sold in the rest of Canada .

Manufacturer/Importer: Eppendorf AG/Eppendorf Canada

Hazard: Potential burn risk when moisture enters the MixMate housing.

Incidents/Injuries: No injuries or property damages reported.

Description/Models: Serial numbers below 08083.

Sold by: Eppendorf Canada, Fisher Scientific Canada, or VWR Canada.

When was distributed: Between 2006 and Dec 2010.

Manufactured in: Germany

Remedy: Please unplug your MixMate 5353 immediately and contact Eppendorf for further information.

Please go to www.eppendorfna.com/mixmaterecall2011 to register your MixMate and initiate the replacement process for a new, improved MixMate version.

Consumer Contact: Eppendorf North America Customer Support, phone: 1-800-645-3050, email: custserv@eppendorf.com.

Remarks: The recalled units are certified to Canadian Standards by UL. For more information about UL's product certification process please visit: www.ul.com



May 5, 2011

